

REV	REQUEST	DESCRIPTION	DATE	DRAWN BY	CHECKED BY
0	BE	DRAWING CREATION	22 MAR 89	F.BERTHET	
A	BE	UPDATE	16 JUI 90	F.BERTHET	
B	1811	SEE REVISIONS NOTE	14 AUG 91	F.BERTHET	
C	40M18694	SEE REVISIONS NOTE	13 APR 11	P.GERARD	E.VERLINDE


MAXIMUM PRACTICAL CONSUMPTION
(99,99 % OF REGULATORS BASED
ON STATISTICAL MEASUREMENTS)

SETTING UPPER LIMIT

AVERAGE PRACTICAL CONSUMPTION
(STUDY BASED ON 20 REGULATORS)

REVISIONS NOTE

REV B: ADDED MAXIMUM PRACTICAL CONSUMPTION AND AVERAGE
PRACTICAL CONSUMPTION CURVES.
REV C: ADDED RMC2018 AND RMC 2022 REGULATORS.



MINIMUM FAA
ISO 122 mm Hg O2 TRACHEAL

OBSERVATIONS :

- EQUIVALENT TO O2 UPPER LIMIT OF CURVE 441091 .
- INCLUDING RMC 1018, 1022, 2018 AND 2022 REGULATORS .

INTERTECHNIQUE proprietary rights are included in the information disclosed herein. Recipient by accepting this document agrees that neither this document nor the information disclosed herein nor any part thereof shall be reproduced or transferred to other documents or used or disclosed to third parties for manufacturing or for any other purpose except as specifically authorized by INTERTECHNIQUE.

Ech/Scale: N/A	Masse/Weight: kg ± g
 dimensions in mm	TOLERANCES GENERALES N/A GENERAL TOLERANCES Angles/Angles: N/A Filetage/thread: N/A Rugosité/Roughness Ra: N/A
	Designation/Title : CONSUMPTION OF ADDED OXYGEN FOR 20 LPM BTPS VENTILATION
DIV 4: OXYGEN SYSTEMS - PLAISIR - F5341	
FORMAT/SIZE : A3 FOLIO/SHEET: 1/1	No. 441090 REV. C.

CT5

B

A IC:3